

Zhuhai Biori Biotechnology Co., Ltd

珠海宝锐生物科技有限公司

Certificate of Analysis

产品检验报告

Product 产品名称	Pre-treatment kit for mRNA capping rate detection	Lot No. 批号	20260101
Specification 规格	10T	Test Date 检验日期	2026.01.01
Test purpose 请检目的	All items	Expiry date 有效期至	2026.12.31
Standard 请检依据	Enterprise Standard	Storage 储存条件	Storage temperatures vary among components. Please store according to the temperature indicated on the reagent label. 各组分储存温度不同，请按试剂标签标识温度存放。
TEST RESULT			
检验结果			
Content 检验项目	Specification 企业标准	Result 检验结果	
	Labels should be clear and easy to recognize. 标识应清晰，易识别。		
Appearance 外观	All components in the reagent kit should be complete and intact, with no leakage of liquid components and no visible foreign matter. 试剂盒内各组分应齐全、完整，液体组分无渗漏、无可见异物。	Conforms 符合规定	
Volume Check 装量检验	The fill volume of each component shall not be less than the labeled value. 各组分装量应不低于标识值。。	Conforms 符合规定	
Physicochemical Properties of the Solution 溶液理化性质	The density, conductivity, and pH value of each solution in the test kit shall be tested separately and shall comply with the finished product physicochemical property standards. 分别检测试剂盒中各溶液的密度、电导率、pH值，应符合成品理化性质标准。	Conforms 符合规定	
Verification by Urea-PAGE Urea-PAGE 验证	RNA samples and probes were pretreated using this kit, followed by gel electrophoresis for identification. After recovery, the sample exhibited a single band. RNA 样品和探针使用该试剂盒进行检测前处理，处理后跑胶鉴定，回收后样品条带单一。	Conforms 符合规定	
Test conclusion 检验结论	Complying with Enterprise Standard. 本品按企业标准检验，结果符合规定。		

Analyzed by (检验人): 何丽媛

Reviewed by (复核人):

