

Zhuhai Biori Biotechnology Co., Ltd

珠海宝锐生物科技有限公司

Certificate of Analysis

产品检验报告

Product 产品名称	dsRNA Residual Quantitative Detection Kit 2.0	Lot No. 批号	20260101
Specification 规格	96T	Test Date 检验日期	2026.01.01
Test purpose 请检目的	All items	Expiry date 有效期至	2026.06.30
Standard 请检依据	Enterprise Standard	Storage 储存条件	Box 1:-20±5°C. Box 2:2-8°C.
TEST RESULT 检验结果			
Content 检验项目	Specification 企业标准	Result 检验结果	
Appearance 外观	Labels should be clear and easy to recognize. 标识应清晰，易识别。 All components in the reagent kit should be complete and intact, with no leakage of liquid components and no visible foreign matter. 试剂盒内各组分应齐全、完整，液体组分无渗漏、无可见异物。 The vacuum aluminum foil bag should not be swollen or leaking. 真空铝箔袋未膨胀或漏气。	Conforms 符合规定	
Correlation coefficient of the standard curve 标准曲线相关系数	R ² >0.99.	Conforms 符合规定	
Accuracy 准确度	It should be within the range of 80-120%. 应在 80-120%范围内。	Conforms 符合规定	
Detection sensitivity 检测灵敏度	Using the dsRNA reference standard provided in the kit as a positive control for detection, its sensitivity was consistent with the description in the instruction manual. 以试剂盒内 dsRNA 标准品作为阳性对照品进行检测，其灵敏度与说明书中描述一致。	Conforms 符合规定	
Repetitiveness 重复性	CV≤ 10.0%.	Conforms 符合规定	
Inter-assay variation 批间差	CV≤ 15.0%.	Conforms 符合规定	
Test conclusion 检验结论	Complying with Enterprise Standard. 本品按企业标准检验，结果符合规定。		

Analyzed by (检验人): 

Reviewed by (复核人): 

