

Zhuhai Biori Biotechnology Co., Ltd.

珠海宝锐生物科技有限公司

Certificate of Analysis

产品检验报告

Product 产品名称	AcuGenix™ FastAmpli qPCR Master Mix with UNG (Universal ROX)	Lot No. 批号	20250526									
Specification 规格	500 T	Test Date 检验日期	2025.05.29									
Test purpose 请检目的	All Items	Expiry date 有效期至	2027.05.25									
Standard 请检依据	Enterprise Standard	Storage 储存条件	-20±5°C									
<p><b>TEST RESULT</b> 检验结果</p> <table border="1"> <thead> <tr> <th>Content 检验项目</th> <th>Specification 企业标准</th> <th>Results 检验结果</th> </tr> </thead> <tbody> <tr> <td>Appearance 外观</td> <td>Clear liquid, no precipitation, no suspended matter, no flocculent. The packaging is intact, no damage, no leakage, and the batch number is correct. 清澈液体，无沉淀，无悬浮物，无絮状物。包装完好，无破损、无漏液，批号无误。</td> <td>Conforms 符合规定</td> </tr> <tr> <td>Difference between batches 批间差异</td> <td>At least one concentration of the enterprise quantitative standard shall be tested for both the reference substance and test sample; there shall be no significant difference in Ct values between groups (difference &lt; 0.5). 对照品与待检品检测至少一个浓度企业定量标准品，组间 Ct 值应无显著性差异 (&lt;0.5)。</td> <td>Conforms 符合规定</td> </tr> </tbody> </table>				Content 检验项目	Specification 企业标准	Results 检验结果	Appearance 外观	Clear liquid, no precipitation, no suspended matter, no flocculent. The packaging is intact, no damage, no leakage, and the batch number is correct. 清澈液体，无沉淀，无悬浮物，无絮状物。包装完好，无破损、无漏液，批号无误。	Conforms 符合规定	Difference between batches 批间差异	At least one concentration of the enterprise quantitative standard shall be tested for both the reference substance and test sample; there shall be no significant difference in Ct values between groups (difference < 0.5). 对照品与待检品检测至少一个浓度企业定量标准品，组间 Ct 值应无显著性差异 (<0.5)。	Conforms 符合规定
Content 检验项目	Specification 企业标准	Results 检验结果										
Appearance 外观	Clear liquid, no precipitation, no suspended matter, no flocculent. The packaging is intact, no damage, no leakage, and the batch number is correct. 清澈液体，无沉淀，无悬浮物，无絮状物。包装完好，无破损、无漏液，批号无误。	Conforms 符合规定										
Difference between batches 批间差异	At least one concentration of the enterprise quantitative standard shall be tested for both the reference substance and test sample; there shall be no significant difference in Ct values between groups (difference < 0.5). 对照品与待检品检测至少一个浓度企业定量标准品，组间 Ct 值应无显著性差异 (<0.5)。	Conforms 符合规定										
Test conclusion 检验结论	Complying with Enterprise Standard 本品按企业标准检验，结果符合规定。											

Analyzed by (检验人):



Reviewed by (复核人):

