

Zhuhai Biori Biotechnology Co., Ltd

珠海宝锐生物科技有限公司

Certificate of Analysis

产品检验报告

| | | | |
|------------------------------------|---|----------------------------|------------|
| Product 产品名称 | 10 × Fast Transfer Buffer | Lot No. 批号 | 20260107 |
| Specification 规格 | 500 mL | Test Date 检验日期 | 2026.01.07 |
| Test purpose 请检目的 | All items | Expiry Date 有效期至 | 2027.01.06 |
| Standard 请检依据 | Enterprise Standard | Storage 储存条件 | 2-8°C |
| TEST RESULT 检验结果 | | | |
| Content 检验项目 | Specification 企业标准 | Results 检验结果 | |
| Appearance 外观 | Clear liquid, no precipitation, no suspended matter, no flocculent. The packaging is intact, no damage, no leakage, and the batch number is correct. 清澈液体，无沉淀，无悬浮物，无絮状物。包装完好，无破损、无漏液，批号无误。 | Conforms 符合规定 | |
| Filling Volume 装量 | Not less than the labeled volume. 组分装量不低于标示值。 | Conforms 符合规定 | |
| Solution Characteristics 溶液外观性状 | Colorless and transparent solution.. 无色透明溶液。 | Conforms 符合规定 | |
| Functional Test 功能性检测 | Diluted to 1× working solution, transferred at 400-500 mA for 20-40 min, good protein transfer efficiency, clear bands on the membrane with no drifting or deformation, and no abnormal heat generation during transfer. 取本试剂稀释至 1×工作液，在 400-500 mA 电流下，转膜 20~40min，蛋白转印效率良好，膜面条带清晰、无漂移变形，转膜过程无异常产热。 | Conforms 符合规定 | |
| Test conclusion 检验结论 | Complying with Enterprise Standard. 本品按企业标准检验，结果符合规定。 | | |

Analyzed by (检验人): 余立韬

Reviewed by (复核人): 张立韬

