

Zhuhai Biori Biotechnology Co.,Ltd

珠海宝锐生物科技有限公司

Certificate of Analysis

产品检验报告

Product 产品名称	ECL Chemiluminescent Substrate Kit	Lot No. 批号	20260108
Specification 规格	100mL/box	Test Date 检验日期	2026.01.10
Test purpose 请检目的	All items	Expiry date 有效期至	2027.01.07
Standard 请检依据	Enterprise Standard	Storage 储存条件	2-8°C
TEST RESULT 检验结果			
Content 检验项目	Specification 企业标准	Results 检验结果	
Appearance 外观	Clear and easily identifiable labeling. All components in the kit are complete and intact, liquid components are free from leakage and visible foreign matter, and the batch number is correct. 标识清晰，易识别。试剂盒内各组分齐全、完整，液体组分无渗漏、无可见异物，批号无误。	Conforms 符合规定	
Volume Inspection 装量检验	The volume of each component should not be less than the labeled value. 各组分装量应不低于标识值。	Conforms 符合规定	
Functional Test 功能性检测	Apply gradient-diluted HRP-containing samples onto PVDF or NC membranes, then add BR4D602 working solution to develop clear gradient results. 将梯度稀释的含 HRP 的样品上样到 PVDF 或 NC 膜上，然后加入 BR4D602 工作液，显影出清晰的梯度结果。	Conforms 符合规定	
Test conclusion 检验结论	Complying with Enterprise Standard. 本品按企业标准检验，结果符合规定。		

Analyzed by (检验人): 张丽

Reviewed by (复核人): 张清

BR4D602-01