

Zhuhai Biori Biotechnology Co.,Ltd

珠海宝锐生物科技有限公司

Certificate of Analysis

产品检验报告

Product 产品名称	HEK293 residual DNA detection kit (qPCR) HEK293 残留 DNA 检测试剂盒 (PCR-荧光探针法)	Lot No. 批号	20260414
Specification 规格	100 Reactions	Test Date 检验日期	2026.04.20
Test purpose 请检目的	All items	Expiry date 有效期至	2028.04.12
Standard 请检依据	Enterprise Standard	Storage 储存条件	≤-20°C
TEST RESULT 检验结果			
Content 检验项目	Specification 企业标准	Results 检验结果	
Appearance 外观检验	Clear liquid, no precipitation, no suspended matter, no flocculent. The packaging is intact, no damage, no leakage, and the batch number is correct. 各液体组分应无渗漏且密封完好无破损。	Conforms 符合规定	
Filling Volume 装量检验	The volume of each composition should meet the minimum requirements as specified in the technical guidelines. HEK293 qPCR MIX 应不低于 1.0mL/管, 阴性对照应不低于 1.0mL/管, HEK293 内标应不低于 1.0mL/管, HEK293 校准品 ST1-ST5 应不低于 0.5mL/管。	Conforms 符合规定	
Negative Concordance Rate 阴性符合率	The negative control should be tested three times with the detection reagent kit, and the results should be negative. 检测 HEK293 阴性对照, 阴性符合率 (-/-) 为 3/3。	Conforms 符合规定	
Precision 精密度	The coefficient of variation (CV%) for low-value precision should be ≤20%; for high-value precision, it should be ≤15%. 检测 HEK293 准确度/精密度参考品 Z1-Z4, Z1 的 CV 值应≤20%, Z2-Z4 的 CV 值应≤15%。	Conforms 符合规定	
Accuracy 准确度	The deviation in low-value accuracy should be ≤ 20%; the deviation in high-value accuracy should be ≤15%. 检测 HEK293 准确度/精密度参考品 Z1-Z4, Z1 的准确度偏差应≤20%, Z2-Z4 的准确度偏差应≤15%。	Conforms 符合规定	
Amplification Efficiency 扩增效率	For calibration standards ST1-ST5 the amplification efficiency of the detection reagent kit should be between 95% and 105%, the slope should be in the range of -3.2 to -3.45. 检测试剂盒校准品 ST1-ST5 其扩增效率应在 95%-105%之间, 即斜率应在-3.2 至-3.45 范围内。	Conforms 符合规定	
Linear Correlation 线性相关性	For calibrators ST1-ST5 of the detection kit, the linear correlation coefficient $ r \geq 0.99$. 检测试剂盒校准品 ST1-ST5 线性相关系数 $ r $ 应≥0.99。	Conforms 符合规定	
Internal Standard Ct Value 内标出值	Operate in strict accordance with the kit instruction manual, detect the kit internal standard, and the Ct value shall be ≤30. 严格按照试剂盒说明书进行操作, 检测试剂盒内标, Ct 出值≤30。	Conforms 符合规定	
Test conclusion 检验结论	Complying with Enterprise Standard. 本品按企业标准检验, 结果符合规定。		

Analyzed by (检验人): 杨丹芳

Reviewed by (复核人): 赵尊记