

Zhuhai Biori Biotechnology Co., Ltd.
 珠海宝锐生物科技有限公司
 Certificate of Analysis
 产品检验报告

| | | | |
|---------------------------------------|---|---------------------|------------|
| Product 产品名称 | Script Max First-Strand cDNA Synthesis Kit | Lot No. 批号 | 20260116 |
| Specification 规格 | 24 T | Test Date 检验日期 | 2026.01.19 |
| Test purpose 请检目的 | All items | Expiry date 有效期至 | 2027.01.15 |
| Standard 请检依据 | Enterprise Standard | Storage 储存条件 | -20±5°C |
| TEST RESULT 检验结果 | | | |
| Content 检验项目 | Specification 企业标准 | Results 检验结果 | |
| Appearance 外观 | Clear liquid, no precipitation, no suspended matter, no flocculent. The packaging is intact, no damage, no leakage, and the batch number is correct. 清澈液体, 无沉淀, 无悬浮物, 无絮状物。包装完好, 无破损、无漏液, 批号无误。 | Conforms 符合规定 | |
| Sensitivity 灵敏度 | The detection rate of sensitivity standard should be consistent with RS. 待检品与对照品对企业灵敏度标准品的检出率应一致。 | Conforms 符合规定 | |
| Difference between batches 批间差异 | The product to be inspected and RS should has at least one concentration of enterprise quantitative standard substance. There should be no significant difference in the Ct value. 对照品与待检品检测至少一个浓度企业定量标准品, 组间 Ct 值应无显著性差异。 | Conforms 符合规定 | |
| Test conclusion 检验结论 | Complying with Enterprise Standard. 本品按企业标准检验, 结果符合规定。 | | |

Analyzed by (检验人): 张彪

Reviewed by (复核人):

