

Zhuhai Biori Biotechnology Co., Ltd.
 珠海宝锐生物科技有限公司
 Certificate of Analysis
 产品检验报告

Product 产品名称	AmpHifi HS DNA Polymerase -V3	Lot No. 批号	20260116
Specification 规格	100 U	Test Date 检验日期	2026.01.19
Test purpose 请检目的	All items	Expiry date 有效期至	2027.01.15
Standard 请检依据	Enterprise Standard	Storage 储存条件	-20±5°C
TEST RESULT 检验结果			
Content 检验项目	Specification 企业标准	Results 检验结果	
Appearance 外观	Clear liquid, no precipitation, no suspended matter, no flocculent. The packaging is intact, no damage, no leakage, and the batch number is correct. 清澈液体, 无沉淀, 无悬浮物, 无絮状物。包装完好, 无破损、无漏液, 批号无误。	Conforms 符合规定	
Product Concentration 产物产量	Compared to the control, amplification product concentration within 90% ≤ test/control ≤ 110%. 与对照品相比, 扩增产物浓度在 90% ≤ 测试/对照 ≤ 110% 范围内。	Conforms 符合规定	
Product peak profile 产物峰形	Peak profile detected by LabChip GX, consistent between test and control. 使用 LabChip GX 检测峰形, 测试与对照一致。	Conforms 符合规定	
Test conclusion 检验结论	Complying with Enterprise Standard. 本品按企业标准检验, 结果符合规定。		

Analyzed by (检验人): 张彪

Reviewed by (复核人):

