

Zhuhai Biori Biotechnology Co.,Ltd

珠海宝锐生物科技有限公司

Certificate of Analysis

产品检验报告

Product 产品名称	AcuGenix™ FastAmpli Lyo-Ready RT-qPCR Master Mix with UNG (Low DNA)		
Component 组分	5×FastAmpli RT Buffer (DG)	Lot No. 批号	2025121103
	12.5×FastAmpli Part Taq/UNG Mix (NA) (with dNTPs) (DG)		2025121103
	50×FastAmpli Part RTase (NA) (DG)		2025121103
Test purpose 请检目的	All items	Test Date 检验日期	2025. 12. 11
Standard 请检依据	Enterprise Standard	Expiry date 有效期至	2027. 06. 10
Storage conditions 储存条件	-20±5° C		
TEST RESULT 检验结果			
Content 检验项目	Specification 企业标准	Results 检验结果	
Appearance 外观	Clear liquid, no precipitation, no suspended matter, no flocculent. The packaging is intact, no damage, no leakage, and the batch number is correct. 清澈液体，无沉淀，无悬浮物，无絮状物。包装完好，无破损、无漏液，批号无误。	Conforms 符合规定	
Sensitivity 灵敏度	The detection rate of sensitivity standard should be consistent with RS. 待检品与对照品对企业灵敏度标准品的检出率应一致。	Conforms 符合规定	
Difference between batches 批间差异	The product to be inspected and RS should has at least one concentration of enterprise quantitative standard substance. There should be no significant difference in the Ct value. 待检品与对照品检测至少一个浓度企业定量标准品，组间 Ct 值应无显著性差异。	Conforms 符合规定	
Host DNA residues 宿主 DNA 残留	The negative initiation rate was ≤ 1/47, and the CT value was > 35 阴性起线率≤1/47，且 CT 值>35	Conforms 符合规定	
Test conclusion 检验结论	Complying with Enterprise Standard 本品按企业标准检验，结果符合规定。		

Analyzed by (检验人):

闪梦丹

Reviewed by (复核人):

甘茂

FM5144-NA