

## Zhuhai Biori Biotechnology Co.,Ltd

珠海宝锐生物科技有限公司

## Certificate of Analysis

产品检验报告

|                                    |                                                                                                                                                                                                          |                     |              |
|------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------|--------------|
| Product<br>产品名称                    | AcuGenix™ Hot Start RT-LAMP Probe Master Mix-V2                                                                                                                                                          |                     |              |
| Component<br>组分                    | AcuGenix™ Bst Hot Start DNA Polymerase-V2                                                                                                                                                                | Lot No.<br>批号       | 20251211     |
|                                    | 2×RT-LAMP Premix Buffer II                                                                                                                                                                               |                     | 20251211     |
|                                    | RNase HII                                                                                                                                                                                                |                     | 20251211     |
|                                    | Neoscript RTase                                                                                                                                                                                          |                     | 20251211     |
| Test purpose<br>请检目的               | All items                                                                                                                                                                                                | Test Date<br>检验日期   | 2025. 12. 11 |
| Standard<br>请检依据                   | Enterprise Standard                                                                                                                                                                                      | Expiry date<br>有效期至 | 2027. 12. 10 |
| Storage conditions<br>储存条件         | -20±5° C                                                                                                                                                                                                 |                     |              |
| TEST RESULT<br>检验结果                |                                                                                                                                                                                                          |                     |              |
| Content<br>检验项目                    | Specification<br>企业标准                                                                                                                                                                                    | Results<br>检验结果     |              |
| Appearance<br>外观                   | Clear liquid, no precipitation, no suspended matter, no flocculent. The packaging is intact, no damage, no leakage, and the batch number is correct.<br>清澈液体，无沉淀，无悬浮物，无絮状物。包装完好，无破损、无漏液，批号无误。            | Conforms<br>符合规定    |              |
| Sensitivity<br>灵敏度                 | The detection rate of sensitivity standard should be consistent with RS.<br>待检品与对照品对企业灵敏度标准品的检出率应一致。                                                                                                     | Conforms<br>符合规定    |              |
| Difference between batches<br>批间差异 | Test at least one concentration of enterprise quantitative standard for test and control samples; no significant intergroup difference in Ct values (<2).<br>对照品与待检品检测至少一个浓度企业定量标准品，组间 Ct 值应无显著性差异 (<2)。 | Conforms<br>符合规定    |              |
| Test conclusion<br>检验结论            | Complying with Enterprise Standard<br>本品按企业标准检验，结果符合规定。                                                                                                                                                  |                     |              |

Analyzed by (检验人): 杨扬

Reviewed by (复核人):



HW206-P01